


Nabídka: 2023205090-1 Eltra TGA		VUT
Kat. číslo	Popis	Množství Cena/jednotku Total
88100-3001	Eltra TGA Thermostep - automatický termogravimetrický analyzátor - zejména pro určení vlhkosti, těkavých látek a popela v uhlí - pec do 1000 °C - 2 karusely pro 20 kelímků, 20 víček - externí ventilátor - kalibrační závaží	1
		
71015	Počítač	1
88400-0584	Monitor	1
72070	Regulátor kyslíku	1
72080	Regulátor dusíku	1
26063	Keramický kelímek	20
26063	Keramické víčko	20
	Mimořádná zákaznická sleva 10 %	-
	Recyklační příspěvek (2,52 Kč / kg) - 65 kg	1
	Doprava, pojištění a balné	1
	Instalace a zaškolení	1
Výsledná cena bez DPH		912 194 Kč

SOUČÁSTÍ CENY ZAŘÍZENÍ JE DÁLE:

- » Instalace zařízení v potřebném rozsahu
- » Kompletní zaškolení na obsluhu zařízení a přípravu vzorků v potřebném rozsahu
- » Po uplynutí 2 měsíční zkušební lhůty další jednodenní školení obsluhy (na vyžádání)
- » Navazující nepřetržitá nezaplatněná podpora našich specialistů z České republiky a Německa (jako vypracování optimálních metod přípravy vzorků, školení).
- » Manuály v českém, anglickém, německém jazyce
- » Prohlášení o shodě v českém jazyce

Nabídka: 2023205090-1 Eltra TGA

VUT

Kat. číslo	Popis	Množství	Cena/jednotku	Total
------------	-------	----------	---------------	-------

100% PODPORA ZÁKAZNÍKŮM:

- » Naše společnost zajišťuje krátké dodací lhůty spotřebních materiálů: 5 – 7 dní.
- » Nabízíme dále roční objednávkové smlouvy, servisní smlouvy, dodávky spotřebních materiálů v pravidelných intervalech.

PRAVIDELNÝ SERVIS – PREVENTIVNÍ PROHLÍDKY

V době trvání záruky i po jejím vypršení zajišťujeme preventivní údržbu námi dodaných zařízení. Tato údržba zahrnuje kontrolu veškerých funkcí zařízení včetně prohlídky (inspekce), čištění, mazání a seřízení mechanických, elektrických, hydraulických a pneumatických částí (dílů). Dále se provádí výměna vadných součástí. Termín vykonání preventivní údržby se dojedná na základě individuální potřeby uživatele.

ZAJIŠTUJEME KVALIFIKOVANÝ SERVIS NA VŠECHNA NÁMI NABÍZENÁ ZAŘÍZENÍ:

- » Záruční a pozáruční servis min. 10 let od zakoupení zařízení
- » Reakční doba servisu do 48 hodin od písemného oznámení při podpisu servisní smlouvy
- » Preventivní prohlídky v pravidelných intervalech
- » Dodávku veškerých náhradních dílů min. 10 let po nákupu zařízení



TGA THERMOSTEP
THERMOGRAVIMETRIC ANALYZER



TGA THERMOSTEP

THERMO- GRAVIMETRIC ANALYSIS

Thermogravimetric analysis is used to determine the mass loss of a sample as a function of the temperature. Suitable instruments include standard laboratory ovens and muffle furnaces with a fixed temperature and subsequent weighing, as well as TGA analyzers with integrated balance and a variable temperature range.

ELTRA's TGA Thermostep combines the drying and ashing process with integrated weighing. For the determination of various thermogravimetric parameters in one analysis cycle, the software allows to define different temperatures and gases (e. g. oxygen or nitrogen) for each analysis step.

RELIABLE AND FLEXIBLE

ELTRA TGA analyzers are an ideal alternative to standard laboratory ovens and muffle furnaces for thermogravimetric analysis. Thanks to a programmable furnace that is connected to an integrated balance, heating and weighing are combined in one instrument. This saves time-consuming manual work and allows for high sample throughput. In addition, typical parameters such as moisture, ash and volatiles can be determined in one analysis run.

The TGA Thermostep processes up to 19 different samples, typically weighing between 500 mg and 5 g, in one analysis cycle. The surrounding atmosphere and temperature of up to 1,000 °C within the


heating chamber can be freely defined by the user during analysis to create a standard operating procedure. The crucible lids, covering of the samples, can be raised or lowered at each stage of the analysis, thus allowing for safe and ASTM-compliant determination of volatiles in coal samples.

TYPICAL SAMPLE MATERIALS

- Coal, coke, secondary fuels, gypsum, flour, plastics, ceramics and many more.

BENEFITS TGA THERMOSTEP

- Measurement of up to 19 samples in one analysis
- Sample weights of up to 5 g
- Fast heating rates, accurate temperature control
- High-performance, precise weighing cell
- Automatic placing and lifting of crucible covers
- Robust design allows for use in laboratories and production



PRECISE RESULTS

HIGH-PERFORMANCE ANALYSIS TECHNOLOGY

The TGA Thermostep is a powerful thermogravimetric analyzer characterized by robust design, high precision and flexibility. It is possible to apply different atmospheres and to use sample weights of up to 5 g. The Thermostep reliably and efficiently measures parameters such as moisture, ash and volatiles according to a user-defined SOP.

PURGING GAS


The TGA Thermostep is very flexible with regards to the purging gas used. At each stage of the analysis either nitrogen, oxygen or surrounding atmosphere can be selected. In the latter, the surrounding atmosphere penetrates into the TGA Thermostep, gently oxidizing the samples.

TEMPERATURE CONTROL

The furnace temperature is monitored by two thermocouples which are not encapsulated. One thermocouple monitors the temperature inside the furnace, the other monitors the temperature within the heating element. Due to the absence of the encapsulation the heating can be controlled quickly and precisely.

NEW: ENCAPSULATED WEIGHING CELL

The latest TGA Thermostep generation features an encapsulated weighing cell with 0.1 mg resolution providing highly precise measurements. The encapsulation isolates the weighing cell from the ambient atmosphere and is extremely stable. The weighing cell is connected to the furnace by a ceramic spindle with pedestal on which the crucibles are placed.



HIGH-CAPACITY HEATING ELEMENTS

The latest TGA Thermostep generation utilizes three heating elements with an improved capacity of 1800 W each (5400 W total power). The result is a faster heating rate and improved stability, especially at high temperatures. The heating elements, located in the upper and lower furnace, provide homogeneous temperature distribution.



COOLING

At the end of each analysis cycle, the cool down process starts. It is possible to program the automatic opening of the TGA furnace lid as a function of the temperature to support the cool down process. For example, the Thermostep can be programmed to open the furnace lid at 650°C halfway and at 500°C completely. In addition, at 300°C an integrated fan is automatically started.

BENEFITS TGA THERMOSTEP

- | Precise measurements
- | Long-term stability
- | Low maintenance
- | Long operating life

Encapsulated weighing cell






Weighing the sample



Option: Crucible lids



Display of analysis results

TGA THERMOSTEP

SIMPLE OPERATION YIELDS QUICK RESULTS

Operation of the TGA Thermostep is simple, convenient and safe. After selecting the Standard Operating Procedure (SOP) in the PC, the sample ID's can be entered into the software. The samples are then weighed in the crucible at the position assigned to the sample ID in the carousel. After one sample has been weighed, the carousel automatically rotates to the next position and the next registered sample can then be weighed in the crucible. Alternatively, a carousel filled with samples which has been weighed externally, can be placed into the analyzer.

It is also possible to position a second carousel with crucible lids above the crucibles. Once the analysis is finished, a new cycle can be started after a short cool-down period.



TGA THERMOSTEP

WORKING IN COMPLIANCE WITH STANDARDS

The ELTRA TGA Thermostep complies with the following international standards, among others:



Norm	Material	Name
D7582 - 12	Coal, coke	Standard Test Methods for Proximate Analysis of Coal and Coke by Macro Thermogravimetric Analysis
D7348 - 08e1	Combustion residues	Standard Test Methods for Loss on Ignition (LOI) of Solid Combustion Residues

THE TGA APPLICATION INSTRUCTIONS

In order to determine thermogravimetric parameters with the TGA Thermostep, an application instruction must be created once. For this purpose, the general conditions for the individual analysis steps are defined once in the Thermostep software. An application for complete coal analysis consists, for example, of the determination of moisture, volatile components and ash. An analysis step includes the specification of start and end temperature, the purging gas to be used, the heating rate and the end criterion.

Both time and mass stability can be selected as criteria for the end of an analysis step. In addition, it can be specified in each analysis step whether the crucible lids are to be put on.



THE ELTRA APPLICATION LABORATORY

For many applications (e.g. TGA analysis of plastics) there are no standards for automated thermogravimetric analysis. However, in order to guarantee a safe and reliable measurement, the ELTRA laboratory in Haan is available for application advice and free trial measurement using the complete analyzer range (TGA, as well as C/S and O/N/H analysis).

Our participation in round robin tests (e.g. ASTM Powder Metallurgy) and in the certification of reference materials (e.g. ECRM 268-1; ECRM 049-1) ensure a consistently high analysis quality.

TGA THERMOSTEP

INTELLIGENT CRUCIBLE MANAGEMENT

SAMPLE CAROUSEL AND REFERENCE CRUCIBLE

The metal sample carousel accepts up to 19 ceramic crucibles. Position no. 20 is reserved for the reference crucible which is part of every measurement. It is used to compensate for weight loss in the crucible, a physical effect which could lead to measurement errors at high temperatures.



Sample carousel and sample weighing



Weighing the crucible



PC-controlled application of crucible lids / crucibles open



PC-controlled application of crucible lids / crucibles closed

SAMPLE WEIGHING

The samples are weighed automatically in the TGA Thermostep. The analyzer allocates the positions of the crucibles in accordance with the number of samples to be measured to ensure the best possible stability during weighing. The software then connects to every occupied position and weighs one sample after the other.

Optionally, an external weighing station is available. Thus it is possible, for example during the final stages of cooling down of the TGA Thermostep, to weigh in a new sample carousel and introduce it to the analyzer with one single movement. This procedure helps to reduce waiting times between two analysis cycles.

CRUCIBLE LIDS

For applications such as the precise and ASTM-compliant analysis of volatiles in coal or of very reactive sample materials, it is essential to cover the crucibles. The TGA Thermostep is equipped not only with a sample carousel but with a second carousel for the crucible lids.

A software-controlled mechanism integrated in the carousel holder lifts and lowers the lids without interrupting the analysis by opening the TGA.

BENEFITS

- | A maximum of 19 samples plus 1 reference crucible
- | Automatic, integrated weighing
- | Separate carousel for crucible lids

CLEARLY STRUCTURED AND CONVENIENT


PC CONTROL WITH WINDOWS®-BASED SOFTWARE

ELTRA's instrument software ensures convenient control and operation of the analyzers. It is multilingual, easy to understand and provides the following features:

- | Custom layouts: user-defined display of windows and storage of different layouts
- | User profiles with multi-level access: creation of different hierarchy levels with different authorizations
- | Storage of analysis results in data base:
- | The data of each analysis is stored and can be called up later
- | Graphic display of temperature profile and mass loss
- | Individual, customer-specific calculations based on the raw data
- | Retrieval of sample-related information from any given time during analysis
- | LIMS communication and data export
- | Applications memory and display of maintenance intervals: individual configuration of maintenance intervals
- | Extensive diagnostics function


CUSTOMIZED VISUALIZATION OF MEASUREMENT RESULTS

- | Display of measurement results after each analysis step
- | Individual calculations possible
- | Ash content can refer to dry or moist samples
- | Export and printing of measurement results



Simultaneous display of temperature (red) and loss in weight (green)

Display of analysis results in groups




Simultaneous display of result, graph and current measurement position

APPLICATIONS


ANALYSIS OF COAL

The determination of moisture, ash and volatiles in coal is a routine application in coal-fired power plants. This can be done manually with various muffle furnaces or in a TGA Thermostep. The optional automated crucible lid management of the Thermostep ensures reliable determination of volatile components. In contrast to analyzers of other manufacturers, it is not necessary to open the Thermostep or run a second analysis cycle to determine the volatiles.



COAL CALIBRATION

Parameters	Mean value	Standard deviation
Moisture	0.32 %	0.08
Ash	6.6 %	0.05
Volatiles	9.1 %	0.3



ANALYSIS OF CHEMICALS

The TGA Thermostep is ideally suited to determine the various degrees of decomposition of chemicals at different temperatures.

The example shows calcium oxalate; the moisture content was analyzed at 105°C, the mass loss at 200°C, 450°C and 850°C.



TYPICAL RESULTS CALCIUM OXALATE

Number of samples

10 samples


Average weight

500 g

Analysis time

2.5 hours

Temperature	Mean mass loss	Standard deviation
105 °C (moisture)	0.2 %	0.01
200 °C	12.2 %	0.02
450 °C	18.9 %	0.05
850 °C	29.8 %	0.03



ANALYSIS OF CEMENT

The **LOI test (loss on ignition)** is particularly important for inorganic materials. For this test the sample is quickly heated to a defined high temperature. This method is used to rapidly determine the volatile components without modifying the sample characteristics too much.

To determine residual moisture in cement an intermediate step at 105°C was added to the LOI test at 1,000°C. The total analysis time for both parameters in a 1 g sample was 70 minutes.



TYPICAL RESULTS CEMENT

Number of samples

10 samples


Average weight

1 g

Analysis time

70 minutes

Parameters	Cement 1	Cement 2
Moisture (105 °C)	0.07 ±0.01 %	3.0 ±0.02 %
LOI (1,000 °C)	0.08 ±0.01 %	1.9 ±0.01 %



TECHNICAL DATA

Sample weight	up to 5 g
Number of samples	19 (+ 1 reference sample)
Number of sample carousels	2 (crucibles and lids)
Precision	0.02 %
Resolution of balance	0.1 mg
Furnace temperature	From room temperature to 1,000 °C
Temperature control	Precision: 2 % or ±2 °C / stability: 2 % or ±2 °C
Gas flow rate	Adjustable from 1 to 10 l/min
Gas pressure	Air 5–6 bar (75–90 psi) / nitrogen 2–4 bar (30–60 psi) / oxygen 2–4 bar (30–60 psi)
Gas purity	Compressed air 99.5 % (oil and fat free) / nitrogen (99.9 %); oxygen (99.9 %)
Operating temperature / humidity	10–35 °C / 20–80 % humidity (not condensating)
Exhaust air	Connection to required / fan included in delivery scope / 4 m ³ per minute / diameter of: 100 mm
Power supply	230 V (±10 %) / one phase / 50/60 Hz / 32 A (analyzer) 230 V (±10 %) / one phase / 50/60 Hz / 2 A (PC, fan)
Weight	65 kg
Dimensions (B x H x T)	55 x 52 x 62 cm
Interfaces	serial and USB
Accessories	Computer, monitor, printer (exact specifications on request)

Eltra GmbH

Retsch-Allee 1-5
42781 Haan
Germany

Telephone: +49 2104 2333-400
Fax: +49 2104 2333-499

info@eltra.com www.eltra.com

VERDER scientific

VERDER SCIENTIFIC

SCIENCE FOR SOLIDS

Verder Scientific is a business field belonging to the Verder Group and sets standards in the development, manufacture and sale of laboratory and analytics devices. Used in quality control, research and development for test-piece preparation and the analysis of solids.

For several decades our companies have supplied production plants and research institutes, laboratories for quality testing and analytics, all kinds of technical specialists and scientists with modern, reliable devices to solve the many and varied challenges they face.

